

# Certificate Of Processing

Prepared for **QUANTUM TATTOO INK LLC**



Gamma Process Run ID **119731A**

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
Tattoo Ink	Lot # 0002	66	CS

Processing Run Start Date/Time:	28-Dec-2019 12:22:18 am	Approx. Downtime (hours):	0.23
Processing Run End Date/Time:	28-Dec-2019 06:25:44 am		

<b>Minimum Specified Dose (kGy):</b>	<b>25.0</b>	<b>Minimum Delivered Dose (kGy):</b>	<b>32.5</b>
<b>Maximum Specified Dose (kGy):</b>	<b>50.0</b>	<b>Maximum Delivered Dose (kGy):</b>	<b>49.8</b>

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest



Reviewed and E-Signed By  
**Garcia, Adrianna (QSRC Technician I)**

Document Content Revision: 1

Signed On 12/29/2019 at 1:52 PM  
UTC / GMT Offset (hh:mm): -8:00

### Processing Location:

STERIS  
1000 South Sarah Place  
Ontario, CA 91761  
Phone: 909-390-9942  
Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.